


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5-C-F*

S.N.: *6986*

Date: *MAR 31 1994*



- a) Output signal amplitude:
0 to +28 mV
- b) Pulse width: *(to 3-2)*
5 to 100 ns
- c) Rise time:
≤ 3 ns
- d) Fall time:
≤ 4 ns
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

R_i = 3 Ω
100 V/400 p.u.
PRF = 2.0 kHz

120/240 V
50-60 Hz
[Signature]