


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-155C-C-N-M-AI-PWT-OT

S.N.: 6993

Date: MAY 19 84



a) Output signal amplitude:

PULSE : 0 TO -2.0 AMP

DC : 0 TO -0.5 AMP

b) Pulse width:
R_L = 10Ω 100 NS TO 10 μS

c) Rise time:
-1 AMP/DIV ≤ 50 NS

d) Fall time:
-1 AMP/DIV ≤ 50 NS

e) PRF: 0 TO 1.0 MHz
-1 AMP/DIV (100% DUTY CYCLE)

f) Jitter, stability:

DC = 0.5 AMP OK

g) Prime power:

120/240 V

50-60 Hz

[Signature]