


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AN-1402-C-F*

S.N.: *6995*

Date: *MAY 3 1994*



a) Output signal amplitude:
0 TO +20 VOLTS

b) Pulse width:
0.3 TO 2.0 NS

c) Rise time:
≤ 100 PS

d) Fall time:
≤ 200 PS

e) PRF:
0 TO 1.0 MHz

f) Jitter, stability:
OK

g) Prime power:
120 / 240 VOLTS
50 + 60 Hz

40 db ATTEN
10 VOLTS/DIV
VMF ≈ 100KHz

[Signature]