


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-P-0T*

S.N.: *7001*

Date: *APRIL 13 1994*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
 $\le 100\text{ PS}$
- d) Fall time:
 $\le 135\text{ PS}$
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK

40 db VATTEN
10 VOLTS PIV
V/F $\approx 10\text{ MHz}$

- g) Prime power:
120/240 VOLTS
*50 * 60 Hz*
- [Signature]*