



PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-1000-C
 S.N.: 7012
 Date: JUNE 20 1994



- a) Output signal amplitude:
0 TO ±10V, TTL + ECL.
- b) Pulse width:
10 NS TO 1 MS
- c) Rise time:
5 NS TO 1 MS
- d) Fall time:
5 NS TO 1 MS
- e) PRF:
1 KHz TO 50 MHz
- f) Jitter, stability:
0%
- g) Prime power:
120/240 V
50-60 Hz

① AVH-5-1-P LITTER TEST
 40dB ATTEN. ∴ 10 VOLTS/DIV
 PRF = 300 KHz
 50 PS/DIV



TEST 54) DELIM.
 7511 IN 100V
 7TU RANGE.

SIMILAR LOW LITTER RESULT FOR MAIN OUT (+5V) BUT SHUNT INPUT TO AVH WITH STD.