


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-PS-P-PN*

S.N.: *7029*

Date: *APRIL 22 1994*



- a) Output signal amplitude:
0 TO 15 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
 ≤ 100 PS
- d) Fall time:
 ≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK

40 dB ATTEN
.5 VOLTS/DIV
PRF = 10 MHz

- g) Prime power:
120/240 VOLTS
50-60 Hz