


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-14.1-P

S.N.: 7036

Date: JUNE 13 1994



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
≤ 10ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN

100 VOLTS/DIV

PRF = 50 KHz

+15 VDC

200 mA
(MAX)