


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-P*

S.N.: *7038*

Date: *JUNE 15 1994*



60 dB ATTEN

100 VERT/DIV

- a) Output signal amplitude:
0 TO +250 V
- b) Pulse width:
≤ 4 nS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC

250 mA (MAX)