


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV6-1-P

S.N.: 7039

Date: JUNE 15 1994



- a) Output signal amplitude:
0 TO +250 V
- b) Pulse width:
≤ 4 nS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
250 mA (MAX)

60 db ATTEN
100 VOLT/DIV

[Signature]