


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5-1-P
S.N.: 7043
Date: JUNE 13 1994



- a) Output signal amplitude:
0 TO +10V.
- b) Pulse width:
≤ 30 PS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

TOP 1.0 NS/DIV
BOT 100 PS/DIV
BOTH 40 dB ATTEN
... 10 VOLTS/DIV
PRF ≈ 600 kHz

+15 VOLTS (DC)
≈ 150 mA (MAX)

[Signature]