


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVI-V-HV2-P-CA-EW-CEA*

S.N.: *7048*

Date: *JUN 2 1994*



60 db ATTEN

100 VOLTS/DIV

PRF = 20 kHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
OK

+ 15 VOLTS DC

250 mA (max)

MS