


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APR-3-PW-C-P-AT-PWT*

S.N.: *7058*

Date: *JULY 29 94*



Ⓐ $R_L = 50\Omega$, PRF = 2 kHz



Ⓑ $R_L = 50\Omega$, PRF = 50 kHz

- a) Output signal amplitude: *0 TO + 200 VOLT*
- b) Pulse width: *R_L 7.50 OHMS
0.1 TO 100 us*
- c) Rise time: *(0.5% MAX
DUTY CYCLE)
≤ 10 ns*
- d) Fall time: *≤ 10 ns*
- e) PRF: *0 TO 10 kHz
(0.5% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
50-60 Hz*

[Signature]