



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-15.3-C*
 S.N.: *7063*
 Date: *JUNE 23 1994*



- a) Output signal amplitude: $0 \text{ TO } \pm 30 \text{ VOLTS (SINE, SP) (+TK1)}$
- b) Pulse width: $\pm 10 \text{ V PC}$
- c) Rise time: $\leq 50 \text{ NS}$
- d) Fall time: $\leq 50 \text{ NS}$
- e) PRF: $0 \text{ TO } 10 \text{ MHz}$
- f) Jitter, stability: *OK*
- g) Prime power:

(A) *500 KHz SINE WITH
10 VOLT OFFSET. RL=50Ω*



120/240 V
50-60 kHz

(B) *500 KHz SQUARE
0 V OFFSET*