


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-C-PN-M-T-TT-0P2-R3*

S.N.: *7074*

Date: *FEB 27 1995*



- a) Output signal amplitude:  
*0 TO ± 350 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 320 VOLTS/DIV  
(70 dB ATTEN)  
PRF ≈ 2 KHz*

*120/240 V  
50 → 60 Hz.  
A*