


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-1A-PS PN*

S.N.: *7077*

Date: *SEPT 20 1994*

- a) Output signal amplitude:  
*0 TO ± 5 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*< 100 PS*
- d) Fall time:  
*< 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 VOLTS*  
*50-60 HZ*



*40 dB ATTEN*  
*5 VOLTS / DIV.*  
*20P 5 NS / DIV*  
*100P 200 PS / DIV*  
*PRF = 1.0 MHz*

*120/240 VOLTS*  
*50-60 HZ*  
*A*