


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3-P-M3*

S.N.: *7088*

Date: *DEC 13 1994*



a) Output signal amplitude:
0 TO +30 VOLTS

b) Pulse width:
2 TO 10 NS

c) Rise time:
≤ 1 NS

d) Fall time:
≤ 1 NS

e) PRF:
0 TO 20 MHz


f) Jitter, stability:
OK

g) Prime power:

*1) +36 VDC
400 mA (MAX.)*

*2) +15 VDC
400 mA (MAX.)*

*Ⓐ 50 dB ATTEN. ∴ 32 VOLTS/DIV
PRF = 20 MHz.*



*Ⓑ AS Ⓐ BUT PRF ≈ 7 MHz
45 NS/DIV.*