


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANM-2-C-P

S.N.: 7089

Date: DEC 15 1994



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
- 10 VOLTS/DIV
PRF ≈ 12 MHz

170/240 VOLTS
50-60 MHz