


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-S1-P

S.N.: 7091

Date: AUG 31 1994



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: ($R_L = 50 \Omega$)
50 NS TO 10 μ S
- c) Rise time:
 ≤ 2 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF: 0 TO 100 KHz
0.5% MAX SWING CYCLE
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN.

100 VOLTS/DIV

PRF = 10 KHz

+24 VDC.

500 mA (MAX)