


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-51-P*
S.N.: *7092*
Date: *AUG 31 1994*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: ($C_L = 50n$)
50 NS TO 1-00)
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 100 KHz*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC
500 mA (MAX)

60 db ATTEN
100 VOLTS/DIV
PRF ≈ 10 KHz