


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-3A-PS-P

S.N.: 7098

Date: JUNE 22 1994



- a) Output signal amplitude: 0 TO +450 VOLTS
- b) Pulse width: (TO 50ns)
≤ 4 ns
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 20 KHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60 Hz

60db ATTEN
= 100 VOLTS/DIV
PRF = 10 KHz