


PULSE GENERATOR

PERFORMANCE CHECK

Model: AN6-3A-C-N

S.N.: 7101

Date: DEC 9 1994



- a) Output signal amplitude: 0 TO 40 VOLTS TO
- b) Pulse width: $R_L \approx 1.0 \Omega$ (40 amperes)
- c) Rise time: 2 μ S TO DC $\leq 0.5 \mu$ S
- d) Fall time: $\leq 0.5 \mu$ S
- e) PRF: 0 TO 1 KHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V 50-60Hz

5 μ S / DIV
10 VOLTS / DIV
 $R_L \approx 1.0 \Omega$
PRF \approx 1 KHz