


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-MV-MV2-C-P*

S.N.: *7100*

Date: *NOV 7 84*



- a) Output signal amplitude:
0 TO +20 V
- b) Pulse width:
0.3 TO 2.0 NS
- c) Rise time:
≤ 100 NS
- d) Fall time:
≤ 200 NS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VAC
50-60 Hz

40 dB ATTEN

10 VOLT/DIV

PRF = 1.0 MHz