


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AK-B3-W-C-P*

S.N.: *7111*


Date: *NOV 10 1994*



Ⓐ $R_L = 50\Omega$, $PRF = 5.0\text{kHz}$

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
0.050 TO 5.045
- c) Rise time:
 $\leq 5\text{NS}$
- d) Fall time:
 $\leq 5\text{NS}$
- e) PRF:
0 TO 5 kHz

- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 Hz



Ⓑ $R_L = 50\Omega$, $PRF = 5\text{kHz}$