


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MV-155A-149MA*

S.N.: *7112*

Date: *SEPT 21 1994*



- a) Output signal amplitude:
0 TO -170V
- b) Pulse width:
20 NS TO DC
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
DC TO 40 MHz
- f) Jitter, stability:
OK
- g) Prime power:
-15 VDC 200mA
+5.8 VDC 40mA



[Signature]