


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-4V2-P-67A-BW-CEA*

S.N.: *7120*

Date: *SEPT 28 1994*



- a) Output signal amplitude:
0 TO +100V
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC.

250 mA (MAX)

60 dB ATTEN

100 VOLTS/DIV

PRF ≈ 20 KHZ