


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARI-V-1A2-P-DT-EN-CEA*

S.N.: *7121*

Date: *SEPT 28 1994*



a) Output signal amplitude:

0 TO 100 V

b) Pulse width:

0 TO 20 NS

c) Rise time:

≤ 2 NS

d) Fall time:

≤ 2 NS

e) PRF:

0 TO 20 KHz

f) Jitter, stability:

OK

g) Prime power:

+ 15 VDC

250 mA (MAX)

60 db ATTEN

100 VOLT/3 DIV

PRF = 20 KHz