


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-MR-P-CA EW CEA*

S.N.: *7122*

Date: *SEPT 28 1994*



- a) Output signal amplitude:
0 TO +100 V.
- b) Pulse width:
0 TO 20 ns
- c) Rise time:
≤ 2 ns
- d) Fall time:
≤ 2 ns
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC.

+15 VDC.

250 mA (MAX)

[Signature]

60 db ATTEN
100 VOLTS DIV
PRF ≈ 20 KHz.