


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVI-U-AV2-P-CA EW-CEA*
S.N.: *7125*
Date: *SEPT 30 1994*



- a) Output signal amplitude:
0 TO 100V
- b) Pulse width:
0 TO 20NS
- c) Rise time:
≤ 2NS
- d) Fall time:
≤ 2NS
- e) PRF:
0 TO 20KHz
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VDC.
250mA (MAX)

60db ATTEN
100V/25NS/PW
PRF = 20KHz

[Signature]