


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-155C-C-N-M-55

S.N.: 7132

Date: NOV 28 1994



- a) Output signal amplitude:
0 TO -2 AMP
- b) Pulse width:
100 NS TO 10 US
- c) Rise time:
≤ 50 NS
- d) Fall time:
≤ 50 NS
- e) PRF: 0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

$R_L = 1.0 \Omega$

TOP: MONITOR OUT (TO HIGH R)
2 VOLTS ∴ 2 AMPS

BOT: VOLTAGE ACROSS $R_L = 1.0 \Omega$