


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P*  
S.N.: *7134*  
Date: *NOV 8 1994*



- a) Output signal amplitude:  
*+25 TO +160 AMP*
- b) Pulse width:  
*≤ 25 ns*
- c) Rise time: *(HAF power)*  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VDC*  
*300 mA MAX*

*≈ 60 AMPs/DIV*  
*PRF = 10 KHz*