


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-PW-C-P-EA*

S.N.: *7136*

Date: *FEB 13 1995*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: ($R_L = 50\Omega$)
50 NS TO 1.0US
- c) Rise time: (*0.15% MAX DUTY CYCLE*)
 $\leq 10\text{ NS}$
- d) Fall time:
 $\leq 10\text{ NS}$
- e) PRF:
0 TO 100 KHZ
- f) Jitter, stability: (*0.5% MAX DUTY CYCLE*)
OK
- g) Prime power:

120/240 VOLTS
50-60 HZ

$R_L = 50\Omega$
PRF = 10 KHZ

[Signature]