


PULSE GENERATOR
PERFORMANCE CHECK

Model: *M-155C-C-PN*


S.N.: *7145*

Date: *OCT 27 1994*



- a) Output signal amplitude:
PULSE: 0 TO ± 2.0 AMP
DC: 0 TO ± 500 mA
- b) Pulse width:
100 NS TO 10 μS
- c) Rise time:
≤ 50 NS
- d) Fall time:
≤ 50 NS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

ⓐ *Test: $R_L = 1\Omega$ PRF = 1 MHz*
≈ 2 AMP PULSE



← 0V
← OFFSET

← PULSE + OFFSET.

120/240 V
50-60 Hz

[Signature]

ⓑ *Noise: $R_L = 1\Omega$ $I_{pulse} = 2.0 AMP$*
 $I_{DC} = 500 mA$