


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMR-G1-C-N-05-CA-R5*

S.N.: *7150*

Date: *DEC 19 1994*



- a) Output signal amplitude:
0 TO -250 VOLTS
- b) Pulse width: *($R_L > 10\Omega$)*
0.1 μ S TO 1.0 MS
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 20 NS
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 VOLTS
50 - 60 Hz

HORIZ: 2 μ S/DIV
VERT: 50 VOLTS/DIV

$R_L = 10$ K
PRF \approx 1 KHz
 $V_{OS} \approx 0$ VOLTS.