


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-143-P-SRCA*

S.N.: *7176*

Date: *NOV 25 94*



- a) Output signal amplitude:  
*0 TO 100 V*
- b) Pulse width:  
*≤ 1 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 600 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VDC*  
*200 mA MAX.*

*60 dB ATTEN*

*-100 VOLTS/DIV*

*PRF = 600 KHz*

*[Signature]*