


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MH-1V3-P-5RCA*

S.N.: *7177*

Date: *NOV 25 1994*



a) Output signal amplitude:  
*0 TO 100 VOLTS*

b) Pulse width:  
*≤ 1 NS*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF:  
*0 TO 600 KHZ*

f) Jitter, stability:  
*OK*

g) Prime power:  
*+15 VDC*  
*200 mA (max)*

*60 dB ATTEN*

*100 VOLTS DIV*

*PRF = 600 KHZ.*

*[Signature]*