


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-HUB-N-SRCA*

S.N.: *719*

Date: *NOV 25 1994*



- a) Output signal amplitude:
0 TO -100V
- b) Pulse width:
≤ 1 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 600 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC.

200mA (MAX)

AB

60 db ATTEN

100 VOLTS/DIV

PRF = 600 KHz