


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-3-P*

S.N.: *7188*

Date: *JAN 23 1995*



- a) Output signal amplitude:
+150 TO 500 VOLTS
- b) Pulse width:
≤ 5 NS (1/2 power)
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN
100 VOLTS/DIV
PRF = 10 KHz

+15 VOLTS
250 mA (max)