


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-A-1-SI-C-N-4H1*
 S.N.: *7196*
 Date: *FEB 2 1995*



- a) Output signal amplitude:
0 TO -300 VOLTS
- b) Pulse width: *TO 50ns*
15 TO 100 NS
- c) Rise time:
≤ 3 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN
100 VOLTS/DIV
PRF = 5 KHz

120/240 VOLTS
50-60 Hz.
[Signature]