


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-3-PW-C-UMB*

S.N.: *7205*

Date: *FEB 20 1995*



- a) Output signal amplitude: *0 TO +25 VOLTS*
- b) Pulse width: *(RL > 6Ω)
300 NS TO 5.0 MS*
- c) Rise time: *(50% MAX DUTY CYCLE)
≤ 50 NS*
- d) Fall time: *≤ 50 NS*
- e) PRF: *0.5 Hz TO 500 Hz
(50% MAX DUTY)*
- f) Jitter, stability: *CLOCK
OK*
- g) Prime power: *120 / 240 V
50 - 60 Hz*

$R_c = 6.0 \Omega$

TOP *500 NS/DIV PRF = 500 Hz*

BOT *500 US/DIV PRF = 50 Hz*