


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-PS-N-TN*

S.N.: *7208*

Date: *FEB 6 1995*



- a) Output signal amplitude:
0 TO -350 V
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

70 dB ATTEN
≈ 320 VOLTS/DIV
PRF = 2 KHz

120/240 VOLTS
5 TO 60 KHz

[Signature]