


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-1AV3-C-M-PN*

S.N.: *7219*

Date: *FEB 21 1995*



a) Output signal amplitude:  
*0 TO 140 VOLTS*

b) Pulse width:  
*0.4 TO 2.0 NS*

c) Rise time:  
*≤ 150 PS*

d) Fall time:  
*≤ 250 PS*

e) PRF: *0 TO 10 MHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*120/240 VOLTS*  
*50-60 Hz*

*50 dB ATTEN*  
*ON 32 VOLTS/DIV*

*[Signature]*