


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-MV-1423-C-P-PN-05-MS41*

S.N.: 7232

Date: JAN 30 1995



- a) Output signal amplitude:
0 TO ± 40 VOLTS
- b) Pulse width:
0.4 TO 2.0 ns
- c) Rise time:
 ≤ 150 ps
- d) Fall time:
 ≤ 250 ps
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 kHz

50 db ATTEN
 ≈ 32 VOLTS / DIV

PRF = 100 kHz

* OS CURRENT RATING
INCREASED TO 400 mA
+ BNC CONNECTOR
PROVIDED FOR OS.