


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AN HV3-C-N-05-MSU1**

S.N.: *7233*

Date: *JAN 30 1995*



- a) Output signal amplitude:
0 TO 40 V_{pk}
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V_{pk}
50 - 60 MHz

50 dB ATTEN
≈ 32 VOLTS/10V
PRF = 100 KHz

* OS current limit
INCREASED TO
400 mA @ BNC
CONNECTORS PROVIDED
FOR OS. *[Signature]*