


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-E3-W-C-T*


S.N.: *7234*

Date: *MAR 6 1995*



- a) Output signal amplitude:
0 TO + 100 VOLTS
- b) Pulse width:
50 NS TO 5.0 NS
- c) Rise time:
≤ 0.5 NS
- d) Fall time:
≤ 1.0 NS
- e) PRF: *0 TO 5 KHZ*
- f) Jitter, stability:
OK
- g) Prime power:
100 ± 10% VOLTS
50-60 HZ

Ⓐ TOP: *10 NS/DIV*
DOT: *1 NS/DIV (RISE TIME)*
60dB ATTEN: *100 VOLTS/DIV*



Ⓑ *20 VOLTS/DIV*
1.0 μS/DIV
R_L = 50 Ω
PRF = 5 KHZ

[Signature]