


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AN-1-C-P*

S.N.: *7235*

Date: *MAR 20 1995*



- a) Output signal amplitude:  
*0 TO +10V.040*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
 *$\le 100\text{PS}$*
- d) Fall time:  
 *$\le 135\text{PS}$*
- e) PRF:  
*0 TO 1.0 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*

*40 db ATTEN*

*10 VOLT/PIV*

*PRF = 300 kHz*

*[Signature]*