


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-AN-HV3-C-N-PN-0T*

S.N.: *7236*

Date: *MAR 16 1995*



- a) Output signal amplitude:
0 TO 740 VOLTS
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

*50 Ω ATTEN
≈ 32 VOLTS/DIV
PRF ≈ 100 KHz.*

*120/240 VOLTS
50-60 Hz.*