


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-C-P-D-M-GT-OT*

S.N.: *7256*

Date: *MAR 15 1995*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS

40 dB ATTEN
10 VOLTS/DIV
PRF ≈ 10 MHz

50-60 MHz
AK