


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-1-C-PN-ET-N*

S.N.: *7258*

Date: *MAR 10 1995*



- a) Output signal amplitude:  
*0 TO ± 250 VOLTS*
- b) Pulse width: *(70 50 ns)*  
*≤ 4 ns*
- c) Rise time:  
*≤ 1 ns*
- d) Fall time:  
*≤ 2 ns*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 VOLTS*  
*50-60 Hz*  
*⊘*

*60 db ATTEN*  
*100 VOLTS/DIV*  
*PRF ≈ 20 KHz*