


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVPP-2-C-P

S.N.: 7262

Date: APRIL 4 1995



- a) Output signal amplitude:
0 TO + 20 VOLTS
- b) Pulse width:
A) 0.4 TO 8.0 NS
B) 8.0 TO 100 NS
- c) Rise time:
 ≤ 200 PS
- d) Fall time:
 ≤ 300 PS
- e) PRF: 0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

TOP RANGE A
1 NS/DIV
10 VOLTS/DIV

BOT RANGE B
10 NS/DIV
10 VOLTS/DIV

90 to 120 volts
50-60 Hz.