


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MVH-MPT-P*

S.N.: *7267*

Date: *APRIL 18 1995*



- a) Output signal amplitude:
- b) Pulse width: *≥ +30 VOLT*
- c) Rise time: *≤ 1 NS*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

40 db ATTEN
10 VOLT DIV
PRF = 300 kHz

+15 VDC
100 mA (MAX)